

ARCS Instrument Replacement Form

Fill out the following form when replacing instruments at an ARCS site:

SV-0709D_cal

System: Skyrad PSPD

Site: Darwin

OLD Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Date of Change (UTC)	Time Start UTC	Time End UTC
8-48	33386	Skyrad	9/11/2006	8.9985	na	10 Sept 07	23:05	11 Sept 07 00:43

NEW Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Change Logger Coefficients and Version
8-48	33376	Skyrad	10/09/07	8.7682	na	X

Replacement completed by: Troy Culgan

NOTES:

Old Part remains on site as spare:

Old part packed for shipment to: SGP

Other old part disposition (explain):

Other comments: